

Manufacturer	Advocate	Roche	Abbott	Home Diagnostics Inc.		LifeScan	Bayer		Arkray
									
	Speaking Redi-Code +[®]	Accu-Chek [®] Compact Plus	Freestyle Lite [™]	TRUEtrack [™]	TRUEresult [™]	OneTouch [®] Ultra2	Ascensia [®] Breeze [®] 2	Ascensia [®] Contour	Reli On [®] Confirm
Audible Results	Yes	No	No	No	No	No	No	No	No
Coding Required	No	No	No	Yes*	No	Yes	No	No	No
Sample Size	1 mg/dL	0.3 mg/dL	0.3 mg/dL	1 mg/dL	0.5 mg/dL	1 mg/dL	1 mg/dL	0.6 mg/dL	0.3 mg/dL
Results Time	5 sec.	5 sec.	5 sec.	10 sec.	4 sec.	5 sec.	5 sec.	5 sec.	7 sec.
AST Approved	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Memory (results)	400	500	400	365	500	500	420	480	360
Daily Alarm	Yes	Yes	Yes	No	Yes	No	No	No	No
Ketone Warning	Yes	No	Yes	No	No	No	No	No	No
Before/After Meal Indicators	Yes	No	No	No	No	No	No	No	No
Hematocrit Range	20 to 60 (%)	15 to 65 (%)	15 to 65 (%)	30 to 55 (%)	20 to 60 (%)	30 to 55 (%)	20 to 55 (%)	0 to 70 (%)	30 to 54 (%)
Battery	(2) AAA	(2) AAA	(1) CR2032/3V	(1) CR2032/3V	(1) CR2032/3V	(1) CR2032/3V	(1) CR2025/3V	(2) CR2032/3V	(1) CR2032/3V
Operating Range	50 to 104 (°F)	50 to 104 (°F)	40 to 104 (°F)	50 to 104 (°F)	50 to 104 (°F)	43 to 111 (°F)	50 to 104 (°F)	41 to 113 (°F)	32 to 122 (°F)
Computer Download	Yes / Free	Yes	Yes	Yes	No	Yes	Yes	Yes	Yes

The information in this comparison chart is based on manufacturer information and Diabetes Health publications. Trademarks are the property of their respective owners.

*Snap-in Code Key